



Air Toxics

Client Sample ID: Y6L0984091617

Lab ID#: 1709470-02A

MODIFIED EPA METHOD TO-14A GC/MS FULL SCAN

| | | | |
|--------------|---------|---------------------|---------------------|
| File Name: | v092809 | Date of Collection: | 9/16/17 11:59:00 PM |
| Dil. Factor: | 1.00 | Date of Analysis: | 9/28/17 07:19 PM |

| Compound | Rpt. Limit (ppbv) | Amount (ppbv) | Rpt. Limit (ug/m3) | Amount (ug/m3) |
|----------------------------------|-------------------|---------------|--------------------|----------------|
| Freon 12 | 0.10 | 0.57 | 0.49 | 2.8 |
| Freon 114 | 0.10 | Not Detected | 0.70 | Not Detected |
| Chloromethane | 0.50 | 0.67 | 1.0 | 1.4 |
| Vinyl Chloride | 0.10 | Not Detected | 0.26 | Not Detected |
| 1,3-Butadiene | 0.10 | 0.048 J | 0.22 | 0.10 J |
| Bromomethane | 0.50 | Not Detected | 1.9 | Not Detected |
| Chloroethane | 0.50 | Not Detected | 1.3 | Not Detected |
| Freon 11 | 0.10 | 0.24 | 0.56 | 1.4 |
| Ethanol | 1.0 | 1.5 | 1.9 | 2.8 |
| Freon 113 | 0.10 | 0.084 J | 0.77 | 0.65 J |
| 1,1-Dichloroethene | 0.10 | Not Detected | 0.40 | Not Detected |
| Acetone | 0.50 | 5.8 | 1.2 | 14 |
| 2-Propanol | 0.50 | 2.8 | 1.2 | 6.9 |
| Carbon Disulfide | 0.50 | Not Detected | 1.6 | Not Detected |
| 3-Chloropropene | 0.50 | Not Detected | 1.6 | Not Detected |
| Methylene Chloride | 0.20 | 0.27 | 0.69 | 0.94 |
| Methyl tert-butyl ether | 0.10 | Not Detected | 0.36 | Not Detected |
| trans-1,2-Dichloroethene | 0.10 | Not Detected | 0.40 | Not Detected |
| Hexane | 0.10 | 0.22 | 0.35 | 0.79 |
| 1,1-Dichloroethane | 0.10 | Not Detected | 0.40 | Not Detected |
| 2-Butanone (Methyl Ethyl Ketone) | 0.50 | 0.70 | 1.5 | 2.0 |
| cis-1,2-Dichloroethene | 0.10 | Not Detected | 0.40 | Not Detected |
| Tetrahydrofuran | 0.50 | Not Detected | 1.5 | Not Detected |
| Chloroform | 0.10 | 0.034 J | 0.49 | 0.16 J |
| 1,1,1-Trichloroethane | 0.10 | Not Detected | 0.54 | Not Detected |
| Cyclohexane | 0.10 | Not Detected | 0.34 | Not Detected |
| Carbon Tetrachloride | 0.10 | 0.11 | 0.63 | 0.69 |
| 2,2,4-Trimethylpentane | 0.50 | 0.52 | 2.3 | 2.4 |
| Benzene | 0.10 | 0.33 | 0.32 | 1.0 |
| 1,2-Dichloroethane | 0.10 | Not Detected | 0.40 | Not Detected |
| Heptane | 0.10 | 0.11 | 0.41 | 0.46 |
| Trichloroethene | 0.10 | Not Detected | 0.54 | Not Detected |
| 1,2-Dichloropropane | 0.10 | Not Detected | 0.46 | Not Detected |
| 1,4-Dioxane | 0.10 | Not Detected | 0.36 | Not Detected |
| Bromodichloromethane | 0.10 | Not Detected | 0.67 | Not Detected |
| cis-1,3-Dichloropropene | 0.10 | Not Detected | 0.45 | Not Detected |
| 4-Methyl-2-pentanone | 0.10 | Not Detected | 0.41 | Not Detected |
| Toluene | 0.10 | 0.63 | 0.38 | 2.4 |
| trans-1,3-Dichloropropene | 0.10 | Not Detected | 0.45 | Not Detected |
| 1,1,2-Trichloroethane | 0.10 | Not Detected | 0.54 | Not Detected |
| Tetrachloroethene | 0.10 | 0.026 J | 0.68 | 0.17 J |
| 2-Hexanone | 0.50 | Not Detected | 2.0 | Not Detected |



Air Toxics

Client Sample ID: Y6L0984091617

Lab ID#: 1709470-02A

MODIFIED EPA METHOD TO-14A GC/MS FULL SCAN

| | | |
|---------------------|----------------|--|
| File Name: | v092809 | Date of Collection: 9/16/17 11:59:00 PM |
| Dil. Factor: | 1.00 | Date of Analysis: 9/28/17 07:19 PM |

| Compound | Rpt. Limit (ppbv) | Amount (ppbv) | Rpt. Limit (ug/m3) | Amount (ug/m3) |
|---------------------------|-------------------|---------------|--------------------|----------------|
| Dibromochloromethane | 0.10 | Not Detected | 0.85 | Not Detected |
| 1,2-Dibromoethane (EDB) | 0.10 | Not Detected | 0.77 | Not Detected |
| Chlorobenzene | 0.10 | Not Detected | 0.46 | Not Detected |
| Ethyl Benzene | 0.10 | 0.087 J | 0.43 | 0.38 J |
| m,p-Xylene | 0.10 | 0.29 | 0.43 | 1.3 |
| o-Xylene | 0.10 | 0.12 | 0.43 | 0.51 |
| Styrene | 0.10 | 0.033 J | 0.42 | 0.14 J |
| Bromoform | 0.10 | Not Detected | 1.0 | Not Detected |
| Cumene | 0.10 | Not Detected | 0.49 | Not Detected |
| 1,1,2,2-Tetrachloroethane | 0.10 | Not Detected | 0.69 | Not Detected |
| Propylbenzene | 0.10 | 0.045 J | 0.49 | 0.22 J |
| 4-Ethyltoluene | 0.10 | 0.11 | 0.49 | 0.54 |
| 1,3,5-Trimethylbenzene | 0.10 | 0.044 J | 0.49 | 0.22 J |
| 1,2,4-Trimethylbenzene | 0.10 | 0.12 | 0.49 | 0.60 |
| 1,3-Dichlorobenzene | 0.10 | Not Detected | 0.60 | Not Detected |
| 1,4-Dichlorobenzene | 0.10 | Not Detected | 0.60 | Not Detected |
| alpha-Chlorotoluene | 0.10 | Not Detected | 0.52 | Not Detected |
| 1,2-Dichlorobenzene | 0.10 | Not Detected | 0.60 | Not Detected |
| 1,2,4-Trichlorobenzene | 0.50 | 0.042 J | 3.7 | 0.31 J |
| Hexachlorobutadiene | 0.50 | Not Detected | 5.3 | Not Detected |
| Butane | 0.50 | 1.2 | 1.2 | 2.8 |
| Decane | 2.0 | Not Detected | 12 | Not Detected |
| Dibromomethane | 0.50 | Not Detected | 3.6 | Not Detected |
| Nonane | 0.50 | Not Detected | 2.6 | Not Detected |
| Octane | 0.50 | Not Detected | 2.3 | Not Detected |
| Pentane | 0.50 | 1.0 | 1.5 | 2.9 |
| Propylene | 1.0 | Not Detected | 1.7 | Not Detected |
| Undecane | 2.0 | Not Detected | 13 | Not Detected |
| Vinyl Acetate | 0.50 | Not Detected | 1.8 | Not Detected |

J = Estimated value.

Container Type: 6 Liter Summa Canister (100% Cert Ambient)

| Surrogates | %Recovery | Method Limits |
|-----------------------|-----------|---------------|
| 1,2-Dichloroethane-d4 | 116 | 70-130 |
| Toluene-d8 | 96 | 70-130 |
| 4-Bromofluorobenzene | 102 | 70-130 |



Air Toxics

Client Sample ID: Y6L0984091617

Lab ID#: 1709470-02B

MODIFIED EPA METHOD TO-14A GC/MS FULL SCAN

| | | | | |
|--------------|----------|---------------------|---------------------|--|
| File Name: | 21092814 | Date of Collection: | 9/16/17 11:59:00 PM | |
| Dil. Factor: | 1.00 | Date of Analysis: | 9/28/17 08:19 PM | |

| Compound | Rpt. Limit (ppbv) | Amount (ppbv) | Rpt. Limit (ug/m3) | Amount (ug/m3) |
|-------------|-------------------|---------------|--------------------|----------------|
| Naphthalene | 0.50 | 0.14 J | 2.6 | 0.76 J |

J = Estimated value.

Container Type: 6 Liter Summa Canister (100% Cert Ambient)

| Surrogates | %Recovery | Method Limits |
|-----------------------|-----------|---------------|
| 1,2-Dichloroethane-d4 | 122 | 70-130 |
| Toluene-d8 | 100 | 70-130 |
| 4-Bromofluorobenzene | 83 | 70-130 |